



PERRY JOHNSON LABORATORY ACCREDITATION, INC.

Certificate of Accreditation

Perry Johnson Laboratory Accreditation, Inc. has assessed the Organization of:

OZ Metrology - Alejandro Oharriz Calderon
Real de Santa Fe No.1201 7, Hacienda Sta Maria Regla, Col. Ciudad del Sol
Querétaro, Querétaro, México. C.P. 76116

*and hereby declares that the Organization is accredited in accordance with
the recognized International Standard:*

ISO/IEC 17025:2017

Whereby, technical competence has been confirmed for the associated scope supplement, in the fields of:

Electrical Calibration
(As detailed in the supplement)

Accreditation claims for conformity assessment activities shall only be made from the addresses referenced within this certificate and shall apply solely to those activities identified in the related scope. This Accreditation is granted subject to the Accreditation Body rules governing the Accreditation referred to above, and the Organization hereby commits to observing and complying with those rules in their entirety.

For PJLA:

Initial Accreditation Date:

Issue Date:

Expiration Date:

August 08, 2026

August 08, 2026

November 30, 2028

Accreditation No.:

Certificate No.:

135790

L26-716

Tracy Szerszen
President

Perry Johnson Laboratory
Accreditation, Inc. (PJLA)
755 W. Big Beaver, Suite 1325
Troy, Michigan 48084

*The validity of this certificate is maintained through ongoing assessments based
on a continuous accreditation cycle. The validity of this certificate should be
confirmed through the PJLA website: www.pjllabs.com*



Certificate of Accreditation: Supplement

OZ Metrology - Alejandro Oharriz Calderon

Real de Santa Fe No.1201 7, Hacienda Sta Maria Regla, Col. Ciudad del Sol

Querétaro, Querétaro, México. C.P. 76116

Contact Name: Alejandro Oharriz Calderon. Phone: 442-741-9737

Accreditation is granted to the facility to perform the following conformity assessment activities:

FIELD OF CALIBRATION	MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	EXPANDED MEASUREMENT UNCERTAINTY ($\pm$) ¹	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED	FLEX CODE	LOCATION OF ACTIVITY
Electrical	Gauss Meter (Magnetic Flux Density)	1 G to 40 G	0.62 % of reading	Power Supply DC BK Precision. Multimeter Keithley 2110, Helmholtz Coils	IEC 61786-1 ISO 10305-2	F1, F2	F, O
Electrical	Gauss Meter (Magnetic Flux Density)	500 G to 10 000 G	0.47 % of reading	Reference Magnets Gaussmeter Lakeshore 410	IEC 61786-1 ISO 10305-2	F1, F2	F, O
Electrical	Magnets, Magnetic Grids and Magnetic Traps (Magnetic Flux Density)	500 G to 10 000 G	0.55 % of reading	Gaussmeter Lakeshore 410	IEC 60404-4 IEC 60404-6 IEC 60404-8-10	F1, F2	F, O

1. The CMC (Calibration and Measurement Capability) is expressed in terms of measurement instrument/aspect being calibrated, range, expanded measurement uncertainty, equipment, and method/procedure. The expanded measurement uncertainty stated for calibrations included on this scope of accreditation represents the smallest measurement uncertainty attainable by the laboratory when performing a more or less routine calibration of a nearly ideal device under nearly ideal conditions. It is typically expressed at a confidence level of 95 % using a coverage factor k (usually equal to 2). The actual measurement uncertainty associated with a specific calibration performed by the laboratory will typically be larger than the measurement uncertainty included on this scope for the same calibration since capability and performance of the device being calibrated and the conditions related to the calibration may reasonably be expected to deviate from ideal to some degree.
2. The laboratory's range of calibration capability for all disciplines for which it is accredited is the interval from the smallest calibrated standard to the largest calibrated standard used in performing the calibration. The low end of this range must be an attainable value for which the laboratory has or has access to the standard referenced. Verification of an indicated value of zero in the absence of a standard is common practice in the procedure for many calibrations but by its definition it does not constitute calibration of zero capacity.



Certificate of Accreditation: Supplement

OZ Metrology - Alejandro Oharriz Calderon

Real de Santa Fe No.1201 7, Hacienda Sta Maria Regla, Col. Ciudad del Sol

Querétaro, Querétaro, México. C.P. 76116

Contact Name: Alejandro Oharriz Calderon. Phone: 442-741-9737

Accreditation is granted to the facility to perform the following conformity assessment activities:

3. Location of activity:

**Location
Code**

Location

- | | |
|---|---|
| F | Conformity assessment activity is performed at the CAB's fixed facility |
| O | Conformity assessment activity is performed onsite at the CAB's customer location |

4. Measurement uncertainties obtained for calibrations performed at customer sites can be expected to be larger than the measurement uncertainties obtained at the laboratory's fixed location for similar calibrations. This is due to the effects of transportation of the standards and equipment and upon environmental conditions at the customer site which are typically not controlled as closely as at the laboratory's fixed location.

5. Flex Codes

F0: When no flexibility is identified. There are no changes to items calibrated, characteristics identified or versions of methods except for updating to the most recent version of a standard method after verification.

F1: The laboratory has the capability to introduce a new instrument, quantity, or gauge for an accredited calibration method.

F2: The laboratory has the capability to introduce the newest revision of an accredited authoritative standard method (with no modifications) identified on the scope

F3: The laboratory has the capability to introduce a new revision of an accredited non-standard method using the same technology or technique identified on the scope

F4: The laboratory has the capability to introduce a validated method that is equivalent to an accredited method (using the same Calibration Equipment or Reference Standards identified on the scope for the same parameter, component, or analyte identified on the line item of the scope.